



Certificate of Traceability

Level I Laboratory Certification

This certifies that LabCorp Central Laboratory, using Bio-Rad Variant II Turbo has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the Diabetes Control and Complications Trial Reference method. The comparison was performed with: University of Missouri SRL#12

The system evaluated was:

Instrument: Variant II Turbo	Calibrator Lot: 64520787, 64520788	Column Lot: 10256L
Reagent Lot: 64550929, 64544803, 64552875	Calibrator Assigned Values: 5.5%, 10.5%	

Date of Certification: May 1, 2024

Certification Expires: May 1, 2025

David South

NGSP Steering Committee Chair

Randee R Little PhD.

NGSP Network Coordinator

Sharon Amulfo

SRL director/ supervisor